

**LIST OF LABORATORY TESTS PERFORMED ON SITE**

SEE REVERSE FOR COUNTING INSTRUCTIONS	AGENCY USE ONLY
CLIA NUMBER:	
STATE LAB NUMBER:	
FACILITY NAME:	
ADDRESS:	
PHONE:	

TEST	INSTRUMENT /KIT	PROFICIENCY TESTING PROGRAM	CPT BILLING CODE	ANNUAL VOLUME	COMPLEXITY	SPECIALTY

COPY IF ADDITIONAL FORMS ARE REQUIRED

SIGNATURE OF REPRESENTATIVE \_\_\_\_\_

DATE \_\_\_\_\_

## GUIDELINES FOR COUNTING TESTS

- For **histocompatibility**, each HLA typing (including disease associated antigens), HLA antibody screen, or HLA crossmatch is counted as one test.
- For **microbiology**, susceptibility testing is counted as one test per group of antibiotics used to determine sensitivity for one organism. Cultures are counted as one per specimen regardless of the extent of identification, number of organisms isolated and number of tests/procedures required for identification.
- For **general immunology**, testing for allergens should be counted as one test per individual allergen.
- For **hematology**, each **measured** individual analyte of a **complete blood count** or **flow cytometry** test that is ordered **and reported** is counted separately. The **wbc differential** is counted as one test.
- For **immunochemistry**, each A/B/O, Rh, antibody screen, crossmatch or antibody identification is counted as one test.
- For **histopathology**, each block (not slide) is counted as one test. Autopsy services are not included. For those laboratories that perform special stains on histology slides, the test volume is determined by adding the number of special stains performed on slides to the total number of specimen blocks prepared by the laboratory.
- For **cytology**, each slide (not case) is counted as one test for both Pap smears and nongynecologic cytology.
- For **clinical cytogenetics**, the number of tests is determined by the number of specimen types processed on each patient; e.g., a bone marrow and a venous blood specimen received on one patient is counted as two tests.
- For **chemistry**, each analyte in a profile counts as one test.
- For **urinalysis**, microscopic and macroscopia examinations, each count as one test. Macroscopics (dipsticks) are counted as one test regardless of the number of reagent pads on the strip.
- For **all specialties/subspecialties**, do not count calculations (e.g., A/g ratiior, MCH, T7, etc.), quality control, quality assurance, or proficiency testing assays.

If you need additional information concerning counting tests for CLIA, please contact your State agency.  
Form